



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**GR2056-BK - Tissue Marking Dye,
Avantik, Black, 2oz, 12/cs (MTO)**

Lot Number: 193148

Expiration Date: 2026-March-31

COMPONENTS

Source: SOL-9001A/193329

Test	Specifications	Results
Color	Black	Black
Viscosity @ 20 - 30 C (cps)	>= 380 and <= 580	399
pH	>= 7.0 and <= 10.0	8.6

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control
This report is electronically produced and is valid without signature.

Form: WM2103

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	James Andrew Dunn	SENIOR PRODUCT MARKETING MANAGER	25-Apr-2024 11:21
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	25-Apr-2024 16:28

* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New_York)