



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

### CERTIFICATE OF ANALYSIS

**RS4021-A - Stain Gill 3 Hematoxylin  
Avantik 4x32 oz (MTO)**

Lot Number: 194713

Expiration Date: 2025-September-30

### COMPONENTS

Source: SOL-5103/194882

Test	Specifications	Results
Color	Purple-Red	Purple-Red
pH @ above 20 C	$\geq 2.05$ and $\leq 2.35$	2.28
Specific Gravity (g/cm <sup>3</sup> )	$\geq 0.98$ and $\leq 1.10$	1.06
UV/VIS Absorbance (A)	$\geq 1.03$ and $\leq 1.23$	1.23
UV/VIS Wavelength (nm)	$\geq 555$ and $\leq 560$	557
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

## Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	James Andrew Dunn	SENIOR PRODUCT MARKETING MANAGER	06-May-2024 08:55
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	09-May-2024 14:47

\* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New\_York)