



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**RS4021-B - Stain Gill 3 Hematoxylin
Avantik 4x1 gal (MTO)**

Lot Number: 196354

Expiration Date: 2025-October-31

COMPONENTS

Source: SOL-5103/196536

Test	Specifications	Results
Color	Purple-Red	Purple-Red
pH @ above 20 C	≥ 2.05 and ≤ 2.35	2.23
Specific Gravity (g/cm ³)	≥ 0.98 and ≤ 1.10	1.04
UV/VIS Absorbance (A)	≥ 1.03 and ≤ 1.23	1.08
UV/VIS Wavelength (nm)	≥ 555 and ≤ 560	557
Functional Test	Pass/Fail	N/A

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.

Form: WM2103

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	James Andrew Dunn	SENIOR PRODUCT MARKETING MANAGER	06-May-2024 09:11
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	09-May-2024 14:50

* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New_York)