



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

### CERTIFICATE OF ANALYSIS

Lot Number: 192134

**RS4245 - Bluing Reagent Classic 4x1  
Gal Avantik (MTO)**

Expiration Date: 2027-February-28

### COMPONENTS

Source: SOL-6600/192315

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	>= 7.80 and <= 8.20	8.20
Specific Gravity (g/cm3)	>= 0.980 and <= 1.015	1.012

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control  
This report is electronically produced and is valid without signature.