



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

Lot Number: 192194

**RS4362-A - Optik Bluing Solution Type1,
Concentrate, 500mL, 4/cs (MTO)**

Expiration Date: 2026-February-28

COMPONENTS

Source: SOL-BU0399/192367

Test	Specifications	Results
pH @ above 22C	≥ 7.90 and ≤ 8.20	8.20
Specific Gravity	Report	1.010

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control
This report is electronically produced and is valid without signature.