



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

### CERTIFICATE OF ANALYSIS

**RS4368-B - Optik Bluing Reagent Type 2,  
1 Gallon (MTO)**

Lot Number: 194724

Expiration Date: 2027-April-30

### COMPONENTS

Source: SOL-6600/194355

Test	Specifications	Results
Color	Clear	Clear
pH @ above 20 C	$\geq 7.80$ and $\leq 8.20$	7.88
Specific Gravity (g/cm <sup>3</sup> )	$\geq 0.980$ and $\leq 1.015$	1.005

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

#### Quality Control

This report is electronically produced and is valid without signature.

Form: WM2103

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

## Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	James Andrew Dunn	SENIOR PRODUCT MARKETING MANAGER	03-May-2024 11:27
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	09-May-2024 14:42

\* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New\_York)