



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

Lot Number: 192738

**RS4396-500 - Bouin's Solution, 500mL
(MTO)**

Expiration Date: 2026-February-28

COMPONENTS

Source: SOL-FXB0U/192898

Test	Specifications	Results
Color	Yellow liquid	Yellow liquid
pH @ 20-30 C	≥ 1.20 and ≤ 2.20	1.56

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control
This report is electronically produced and is valid without signature.