



19 Chapin Road - Building C, Pine Brook, NJ 07058  
800.783.9424 | avantik-us.com

### CERTIFICATE OF ANALYSIS

Lot Number: 193838

**RS4421-500 - Light Green Counterstain -  
500mL (MTO)**

Expiration Date: 2025-May-31

#### COMPONENTS

*Source: SOL-STLGC/193966*

Test	Specifications	Results
Color	Dark green	Dark green
pH @ 20-30 C	>= 2.50 and <= 3.50	3.03
UV/VIS Absorbance (A)	>= 0.10 and <= 0.25	0.16
UV/VIS Wavelength (nm)	>= 630 and <= 635	632

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control

This report is electronically produced and is valid without signature.

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

## Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	James Andrew Dunn	SENIOR PRODUCT MARKETING MANAGER	15-Apr-2024 16:29
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	16-Apr-2024 14:20

\* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New\_York)