



19 Chapin Road - Building C, Pine Brook, NJ 07058  
 800.783.9424 | avantik-us.com

### CERTIFICATE OF ANALYSIS

**RS4429-1GL - Nuclear Fast Red - Gallon (MTO)**

Lot Number: 193285

Expiration Date: 2025-March-31

#### COMPONENTS

Source: SOL-STNFR/193438

Test	Specifications	Results
Color	Dark Red	Dark Red
Specific Gravity	$\geq 0.97$ and $\leq 1.03$	1.00
UV/VIS Absorbance (A)	$\geq 0.10$ and $\leq 0.40$	0.26
UV/VIS Wavelength (nm)	$\geq 508$ and $\leq 516$	513
UV/VIS Absorbance (A)	$\geq 0.10$ and $\leq 0.40$	0.29
UV/VIS Wavelength (nm)	$\geq 540$ and $\leq 550$	545
Functionality	Pass/Fail	Pass

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control  
 This report is electronically produced and is valid without signature.

Form: WM2103

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

## Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	James Andrew Dunn	SENIOR PRODUCT MARKETING MANAGER	25-Apr-2024 11:31
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	25-Apr-2024 16:38

\* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New\_York)