



19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

**RS4545-125 - Acid Fast Bacteria (AFB)
Stain Kit, 125mL (MTO)**

Lot Number: 194046

Expiration Date: 2024-December-31

COMPONENTS

Source: SOL-KC321/194451

| Test | Specifications | Results |
|---|-----------------------------|-------------------------|
| Visual Inspection | Dark Red, Opaque Liquid | Dark Red, Opaque Liquid |
| UV/VIS Spectrophometry Absorbance (A) | ≥ 1.90 and ≤ 2.60 | 2.43 |
| UV/VIS Spectrophometry Wavelength (nm) | ≥ 543 and ≤ 547 | 546 |

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Control
This report is electronically produced and is valid without signature.